


HIT CLIP (Z117A)

Magnethalter und Klettbefestigung
für METRAHIT Multimeter & Kalibratoren mit Gummischutzhülle
Magnetic holder and belt strap
for METRAHIT Multimeter & Calibrators with Rubber Holster

3-349-664-15
2/9.14



Achtung!

Geräte mit Messfunktionsumschaltung durch Relais*

Im Betrieb darf der Magnet nicht in der Parkposition innerhalb der Gummischutzhülle fixiert sein. Es ist mindestens ein Abstand von 5 cm zum Gehäuseboden einzuhalten.



Attention!

Instruments with Measuring Function Switching by Relay*

During operation, the magnet may not be fixed inside the protective rubber cover in its stand-by position. A minimum distance of 5 cm from the housing bottom must be observed.

* METRACAL MC, METRAHIT ENERGY, ULTRA,
METRAHIT 27M, 27I, H+E CAR, 30M:

Der Betrieb dieser Geräte in der Nähe starker Magnete bzw.
Magnetfelder kann die Funktion der Relais im Gerät beeinflussen.
*Operating these instruments in the proximity of strong magnets
and/or magnetic fields may interfere with the function of the relays
in the instrument.*



